

# **Molbio Diagnostics Limited**

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India S+918326724888 www.molbiodiagnostics.com

#### **MOLBIO DIAGNOSTICS LIMITED**

### EN ISO 13485:2016 certified company

### **CERTIFICATE OF ANALYSIS**

License No.: MFG/IVD/2019/000003

Date: 28/09/2025

Doc ID: COA/TP/VUCPK478L

Product Name	Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit				
Generic Name	Sample preparation kit for nucleic acid extraction on Trueprep® AUTO v2				
Lot No.	VUCPK478L				
Mfg. Date	2025-08				
Exp. Date	2027-07				
Pack Size	50T				
Storage condition	2°C to 40°C				

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation				
Α	Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)						
. 1	Universal Reagent Pack	Universal Reagent Pack Should be Present					
2	Universal Cartridge Pack	Should be Present	Present				
3	Disposable Transfer Pipette (3 ml Graduated)	Should be Present	Present				
4	Package Insert	Should be Present	Present				
5	Desiccant pouch	Should be Present	Present				
6	Colour of desiccant pouch	Should be Orange in Colour	Orange Colour				
В	Seal integrity test	Should Pass	Pass				

FM/QT/224-03

Page 1 of 2



# **Molbio Diagnostics Limited**

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India U+918326724888 www.molbiodiagnostics.com

	Performano	e on Tru	elab® Real Time	micro PCR A	nalyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)	
	Specificity Tests	Test	IC	Test 1 (Average)	Test 2 (Average)	IC (Average)
Negat	ive Control					
1	Negative MTB	ND		ND	NA	27.14
2	Negative HBV	ND	Control should be detected	ND	NA	26.96
3	Negative HCV	ND		ND	NA	28.13
4	Negative HIV-1	ND		ND	NA	27.54
5	Negative MTB Plus	ND		ND	NA	27.09
Positi	ve Control			Louisia		
1	Positive MTB	Test and Control shoul be detected		30.32	NA	26.77
2	Positive HIV-1	Test and Control should be detected		31.74	NA	27.60
3	Positive MTB Plus	Test and Control should be detected		31.56	NA	26.46
Non T	emplate Control Test	for COV	ID-19			Access of the second se
1	NTC	Test 1, Test 2 and Control should not be detected		ND	ND	ND

\*Ct-Cycle Threshold Template Control

\*IC- Internal Control

\*ND- Not Detected

\*NA- Not Applicable

\*NTC-Non

The lot is tested and verified as per QC Standards and is approved.

Name: Dr.Praveen Kumar MK **Department:** Quality Control Designation:QC Head

FM/QT/224-03

Page 2 of 2