

# **Molbio Diagnostics Limited**

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa - 403722, India 3+918326724888 www.molbiodiagnostics.com

#### **MOLBIO DIAGNOSTICS LIMITED**

## EN ISO 13485:2016 certified company

### **CERTIFICATE OF ANALYSIS**

License No.: MFG/IVD/2019/000003

Date: 15/10/2025

Doc ID: COA/TP/VUCPK484L

Product Name	Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit					
Generic Name	Sample preparation kit for nucleic acid extraction on Trueprep® AUTO v2					
Lot No.	VUCPK484L					
Mfg. Date	2025-08					
Exp. Date	2027-07					
Pack Size	50T					
Storage condition	2°C to 40°C					

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation					
. A	Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)							
1	Universal Reagent Pack	Universal Reagent Pack Should be Present						
2	Universal Cartridge Pack	Should be Present	Present					
3	Disposable Transfer Pipette (3 ml Graduated)	Should be Present	Present					
4	Package Insert	Should be Present	Present					
5	Desiccant pouch	Should be Present	Present					
6	Colour of desiccant pouch	Should be Orange in Colour	Orange Colour					
В	Seal integrity test	Should Pass	Pass					

FM/QT/224-03

Page 1 of 2



## **Molbio Diagnostics Limited**

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India S+918326724888 www.molbiodiagnostics.com

	Performanc	e on Tru	elab® Real Time	micro PCR A	nalyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)	
	Specificity Tests	Test	IC	Test 1 (Average)	Test 2 (Average)	IC (Average)
Negat	ive Control					
1	Negative MTB	ND	Control should be detected	ND	NA	27.17
2	Negative HBV	ND		ND	NA	27.51
3	Negative HCV	ND		ND	NA	28.20
4	Negative HIV-1	ND		ND	NA	27.21
5	Negative MTB Plus	ND		ND	NA	27.87
Positi	ve Control		*	***************************************		
1	Positive MTB	Test and Control should be detected		30.69	NA	27.40
2	Positive HCV	Test and Control should be detected		30.55	NA	26.75
3	Positive MTB Plus	Test and Control should be detected		29.78	NA	27.79
Non T	emplate Control Test	for COV	ID-19			
1	NTC	Test 1, Test 2 and Control should not be detected		ND	ND	ND

\*Ct-Cycle Threshold \*IC- Internal Control \*NA- Not Applicable \*NTC-Non \*ND- Not Detected Template Control

The lot is tested and verified as per QC Standards and is approved.

Jou Name: Dr. Praveen Kumar MK **Department:** Quality Control Designation:QC Head

FM/QT/224-03

Page 2 of 2