

Molbio Diagnostics Limited

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India

MOLBIO DIAGNOSTICS LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2019/000003

Date: 16/10/2025

Doc ID: COA/TP/VUCPK485L

Product Name	Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit				
Generic Name	Sample preparation kit for nucleic acid extraction on Trueprep® AUTO v2				
Lot No.	VUCPK485L				
Mfg. Date	2025-08				
Exp. Date	2027-07				
Pack Size	50T				
Storage condition	2°C to 40°C				

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation					
Α	Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)							
1	Universal Reagent Pack	Should be Present	ould be Present Present					
2	Universal Cartridge Pack	Should be Present	Present					
3	Disposable Transfer Pipette (3 ml Graduated)	Should be Present	Present					
4	Package Insert	Should be Present	Present					
5	Desiccant pouch	Should be Present	Present					
6	Colour of desiccant pouch	Should be Orange in Colour	Orange Colour					
В	Seal integrity test	Should Pass	Pass					

FM/QT/224-03

Page 1 of 2



Molbio Diagnostics Limited

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India S+91 832 6724888 www.molbiodiagnostics.com

	Performand	e on Tru	elab® Real Time	micro PCR A	nalyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)	
	Specificity Tests	Test	IC	Test 1 (Average)	Test 2 (Average)	IC (Average)
Negat	ive Control					
1	Negative MTB	ND	Control should be detected	ND	NA	27.93
2	Negative HBV	ND		ND	NA	27.14
3	Negative HCV	ND		ND	NA	27.87
4	Negative HIV-1	ND		ND	NA	28.32
5	Negative MTB Plus	ND		ND	NA	27.79
Positi	ve Control					
1	Positive MTB		d Control should e detected	32.32	NA	27.49
2	Positive HCV	Test and Control should be detected		31.02	NA	26.93
3	Positive MTB Plus	Test and Control should be detected		30.85	NA	27.38
Non T	emplate Control Test	for COV	ID-19			
1	NTC	Contro	1, Test 2 and I should not be detected	ND	ND	ND

*Ct-Cycle Threshold *IC- Internal Control *ND- Not Detected *NA- Not Applicable *NTC-Non Template Control

The lot is tested and verified as per QC Standards and is approved.

Name: Dr.Praveen Kumar MK **Department:** Quality Control Designation:QC Head

Page 2 of 2

FM/QT/224-03