

Molbio Diagnostics Limited

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India S+918326724888 www.molbiodiagnostics.com

MOLBIO DIAGNOSTICS LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2019/000003

Date: 17/10/2025

Doc ID: COA/TP/VUCPK486L

Product Name	Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit				
Generic Name Sample preparation kit for nucleic acid extraction on Trueprep®					
Lot No.	VUCPK486L				
Mfg. Date	2025-10				
Exp. Date	2027-09				
Pack Size	50T				
Storage condition	2°C to 40°C				

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Components in each kit Specification						
Α	Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)							
1	Universal Reagent Pack	Should be Present	Present					
2	Universal Cartridge Pack	Should be Present	Present					
3	Disposable Transfer Pipette (3 ml Graduated)	Should be Present	Present					
4	Package Insert	Should be Present	Present					
5	Desiccant pouch	Should be Present	Present					
6	Colour of desiccant pouch	Should be Orange in Colour	Orange Colour					
В	Seal integrity test	Should Pass	Pass					

FM/QT/224-03

Page 1 of 2



Molbio Diagnostics Limited

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa - 403722, India 3+918326724888 www.molbiodiagnostics.com

	Performand	e on Tru	elab® Real Time	micro PCR A	nalyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)	
	Specificity Tests	Test	IC	Test 1 (Average)	Test 2 (Average)	IC (Average)
Negat	ive Control					
1	Negative MTB	ND	Control should be detected	ND	NA	28.10
2	Negative HBV	ND		ND	NA	27.97
3	Negative HCV	ND		ND	NA	26.94
4	Negative HIV-1	ND		ND	NA	27.01
5	Negative MTB Plus	ND		ND	NA	27.74
Positiv	ve Control		•			
1	Positive MTB	Test and Control should be detected		30.14	NA	27.67
2	Positive HCV	Test and Control should be detected		30.25	NA	26.87
3	Positive MTB Plus	Test and Control should be detected		30.49	NA	28.49
Non To	emplate Control Test	for COVI	D-19			
1	NTC	Test 1, Test 2 and Control should not be detected		ND	ND	ND

*Ct-Cycle Threshold *IC- Internal Control *NTC- Non *ND- Not Detected *NA- Not Applicable Template Control

The lot is tested and verified as per QC Standards and is approved.

Name: Dr.Praveen Kumar MK **Department:** Quality Control Designation:QC Head

FM/QT/224-03

Page 2 of 2