

**MOLBIO DIAGNOSTICS LIMITED**  
**EN ISO 13485:2016 certified company**  
**CERTIFICATE OF ANALYSIS**

Licence No.: MFG/IVD/2021/000071

Date: 22/11/2025

Document ID: COA/TN/206TX

Product Name	Truenat <sup>®</sup> MTB-RIF Dx
Generic Name	Chip-based Real Time PCR Test for Rifampicin Resistant <i>Mycobacterium tuberculosis</i>
Lot No.	206TX
Set No.	NA
Mfg. Date	2025-10
Exp. Date	2027-09
Pack Size	5T , 20T
Storage condition	2°C to 30°C

**PRECAUTION: Refer Safety Data Sheet and Instructions for Use.**

Sr. No.	Components in each kit	Specification	Observation
01	Number of pouches as per pack size	Should be present	Present
02	Pouch Contents: 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	Should be present	Present
03	Colour of desiccant pouch	Should be present and Orange in Colour	Present / Orange Colour
04	Package Insert	Should be present	Present
05	Seal integrity test	Should pass	Pass

Performance testing data				
Samples	Ct Values			Results
	Tm		Control	
	P2	P1		
Positive sample 1	66.78 ± 0.37	66.73 ± 0.29	70.01 ± 0.62	RIF Resistance detected
Positive sample 2	72.49 ± 0.21	66.70 ± 0.25	69.69 ± 0.21	RIF Resistance not detected
Positive sample 3	72.53 ± 0.40	67.13 ± 0.32	69.75 ± 0.25	RIF Resistance not detected
Tm- Melting temperature, Ct- Cycle threshold				

\*Mfg- Manufacturing, Exp- Expiry

The lot is tested and verified as per QC standards and is approved.

Name : Dr. Praveen Kumar  
Department: Quality Control  
Designation: QC Head



FM/QT/219-03

Page 1 of 1