

MOLBIO DIAGNOSTICS LIMITED

Molbio Diagnostics Limited

(Formerly Molbio Diagnostics Private Limited)

L-46, Phase II D, Verna Industrial Estate

Verna, Goa - 403722, India

+91 832 6724888 www.molbiodiagnostic.com

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000003

Date: 18/12/2025

Doc ID: COA/TN/PCI057

BRAND NAME	Truenat [®] Positive Control Kit - Panel I		
GENERIC NAME	Control Kit for Truenat [®] Chip-based micro PCR Tests: MTB, MTB Plus, MTB-RIF Dx, Beta CoV, SARS CoV-2, COVID-19 and H1N1		
LOT NO	PCI057		
MFG. DATE	2025-12		
EXP. DATE	2026-11		
PACK SIZE	08T		
STORAGE	2°C to 40°C		A.R. No. : B-PCI-057

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Negative control	Present	Present
2	Positive control (08- strip tubes containing dried down positive control)	Present	Present
3	Package insert	Present	Present

Observation (Ct Value)			
Sr. No.	Truenat [®] test name	Positive control	Negative control
1	MTB	26.83 ± 0.41	ND
2	MTB Plus	25.44 ± 0.34	ND
3	MTB-RIF Dx	RIF Resistance not detected	NA
4	SARS CoV-2	27.49 ± 0.73	ND
5	COVID-19	Orf 1A: 27.13 ± 0.15 E gene: 26.65 ± 0.42	ND
6	H1N1	25.68 ± 0.43	ND

*ND- Not detected, Ct- Cycle threshold, RIF- Rifampicin, NA- Not Applicable

*A deviation of = <1.5 is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS

MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE



QUALITY CONTROL

Name: Dr. Praveen Kumar MK

Designation: QC Head

Page 1 of 1