

**MOLBIO DIAGNOSTICS LIMITED**

**EN ISO 13485:2016 certified company**

**CERTIFICATE OF ANALYSIS**

**License No.:** MFG/IVD/2021/000064

**Date:** 16/01/2026

**Doc ID:** COA/TP/208MSPT

<b>Product Name</b>	Trueprep® AUTO MTB Sample Pre-treatment Pack
<b>Generic Name</b>	Sample pre-treatment pack for MTB samples
<b>Lot No.</b>	208MSPT
<b>Mfg. Date</b>	2026-01
<b>Exp. Date</b>	2027-12
<b>Pack Size</b>	50T
<b>Storage condition</b>	2°C to 40°C

**PRECAUTION:** Refer Safety Data Sheet and follow instructions for use.

<b>Sr. No.</b>	<b>Components in each kit</b>	<b>Specification</b>	<b>Observation</b>
<b>A</b>	<b>Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)</b>	<b>(Present (P)/Not Present (NP))</b>	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present

Performance on Truelab® Real Time micro PCR Analyzer					
Sr. No.	Specificity Tests	Specification		Test Lot	
		Test	Control	Test (Average)	Control (Average)
<b>Negative Control</b>					
1	Negative MTB	ND	Control should be detected	ND	26.17
<b>Positive Control</b>					
1	Positive MTB	Test and Control should be detected		26.22	27.00
2	Positive MTB	Test and Control should be detected		32.01	26.83
<b>Non Template Control</b>					
1	NTC	Test and Control should not be detected		ND	ND

\*Ct- Cycle Threshold    \*IC- Internal Control    \*ND- Not Detected    \*NTC- Non Template Control

**The lot is tested and verified as per QC standards and is approved.**



Name: Dr. Praveen Kumar  
 Department: Quality Control  
 Designation: QC Head