



## Molbio Diagnostics Limited

(Formerly Molbio Diagnostics Private Limited)

L-46, Phase II D, Verna Industrial Estate

Verna, Goa - 403722, India

+91 832 6724888 [www.molbiodiagnostic.com](http://www.molbiodiagnostic.com)

### MOLBIO DIAGNOSTICS LIMITED

EN ISO 13485:2016 certified company

### CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2021/000064

Date: 02/02/2026

Doc ID: COA/TP/209MSPT

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack
Generic Name	Sample pre-treatment pack for MTB samples
Lot No.	209MSPT
Mfg. Date	2026-01
Exp. Date	2027-12
Pack Size	50T
Storage condition	2°C to 40°C


**PRECAUTION:** Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present

Performance on Truelab® Real Time micro PCR Analyzer					
Sr. No.	Specificity Tests	Specification		Test Lot	
		Test	Control	Test (Average)	Control (Average)
<b>Negative Control</b>					
1	Negative MTB	ND	Control should be detected	ND	28.01
<b>Positive Control</b>					
1	Positive MTB	Test and Control should be detected		26.69	26.89
2	Positive MTB	Test and Control should be detected		32.17	27.31
<b>Non Template Control</b>					
1	NTC	Test and Control should not be detected		ND	ND

\*Ct- Cycle Threshold    \*IC- Internal Control    \*ND- Not Detected    \*NTC- Non Template Control

**The lot is tested and verified as per QC standards and is approved.**



Dr. Praveen Kumar  
Name: Dr. Praveen Kumar  
Department: Quality Control  
Designation: QC Head