

**MOLBIO DIAGNOSTICS LIMITED**  
**EN ISO 13485:2016 certified company**  
**CERTIFICATE OF ANALYSIS**

Licence No.: MFG/IVD/2019/000018

Date: 23/12/2025  
Document ID: COA/TN/HH008

Product Name	Truenat <sup>®</sup> H3N2/H1N1
Generic Name	Chip-based Real Time Duplex PCR Test for H3N2 and H1N1
Lot No.	HH008
Set No.	NA
Mfg. Date	2025-09
Exp. Date	2027-08
Pack Size	20T
Storage condition	2°C to 30°C

**PRECAUTION:** Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
01	Number of pouches as per pack size	Should be present	Present
02	Pouches Contents: 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	Should be present	Present
03	Colour of desiccant pouch	Should be present and Orange in Colour	Present / Orange Colour
04	Package Insert	Should be present	Present
05	Seal integrity test	Should pass	Pass

Performance testing data			
Samples	Ct Values		
	H3N2	H1N1	Control
Non Template Control	ND	ND	ND
Negative Sample	ND	ND	27.29 ± 0.69
Positive Sample 1	26.76 ± 1.02	ND	27.35 ± 1.44
Positive Sample 2	ND	30.58 ± 0.47	27.50 ± 1.20
Positive Sample 3	29.10 ± 0.89	29.83 ± 1.48	27.52 ± 0.71

Mfg- Manufacturing, Exp- Expiry

The lot is tested and verified as per QC Standards and is approved.

  
  
**Name : Dr. Praveen Kumar MK**  
**Department: Quality Control**  
**Designation: QC Head**  
*for*