

**MOLBIO DIAGNOSTICS LIMITED**  
EN ISO 13485:2016 certified company  
**CERTIFICATE OF ANALYSIS**

Licence no.: MFG/IVD/2019/000003

Date: 07/04/2026  
Document ID: COA/TN/PCI064

<b>Product Name</b>	Truenat <sup>®</sup> Positive Control Kit - Panel I
<b>Generic Name</b>	Control Kit for Truenat <sup>®</sup> Chip-based micro PCR Tests: MTB, MTB Plus, MTB-RIF Dx, Beta CoV, SARS CoV-2, COVID-19 and H1N1
<b>Lot No.</b>	PCI064
<b>Mfg. Date</b>	2026-03
<b>Exp. Date</b>	2027-02
<b>Pack Size</b>	8T
<b>Storage condition</b>	2°C to 40°C

**PRECAUTION:** Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
01	Negative control	Should be present	Present
02	Positive control (08- strip tubes containing dried down positive control)	Should be present	Present
03	Package insert	Should be present	Present

**MOLBIO DIAGNOSTICS LIMITED**  
**EN ISO 13485:2016 certified company**  
**CERTIFICATE OF ANALYSIS**



Licence no.: MFG/IVD/2019/000003

Date: 07/04/2026  
Document ID: COA/TN/PCI064

Performance testing data			
Sr. No.	Truenat <sup>®</sup> test name	Positive control	Negative control
1	MTB	30.97 ± 0.08	ND
2	MTB Plus	28.16 ± 0.26	ND
3	MTB-RIF Dx	RIF Resistance detected	Indeterminate
4	SARS CoV-2	28.13 ± 0.37	ND
5	COVID-19	Orf 1A: 28.03 ± 0.32 E gene: 29.09 ± 0.12	ND
6	H1N1	27.93 ± 0.63	ND

\*Mfg- Manufacturing, Exp- Expiry

**The lot is tested and verified as per QC standards and is approved.**

  
  
Name: Dr. Praveen Kumar MK  
Department: Quality Control  
Designation: QC Head

FM/QT/284-00

Page 2 of 2