

MOLBIO DIAGNOSTICS LIMITED
EN ISO 13485:2016 certified company
CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2025/000136

Date: 25/05/2026

Doc ID: COA/TP/252MSPT

Product Name	Trueprep [®] AUTO MTB Sample Pre-treatment Pack
Generic Name	Sample pre-treatment pack for MTB samples
Lot No.	252MSPT
Mfg. Date	2026-05
Exp. Date	2028-04
Pack Size	50T
Storage condition	2°C to 40°C


PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep[®] AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present

Performance on Truelab® Real Time micro PCR Analyzer					
Sr. No.	Specificity Tests	Specification		Test Lot	
		Test	Control	Test (Average)	Control (Average)
Negative Control					
1	Negative MTB	ND	Control should be detected	ND	27.45
Positive Control					
1	Positive MTB	Test and Control should be detected		26.45	26.31
2	Positive MTB	Test and Control should be detected		30.73	27.03
Non Template Control					
1	NTC	Test and Control should not be detected		ND	ND

*Ct- Cycle Threshold *IC- Internal Control *ND- Not Detected *NTC- Non Template Control

The lot is tested and verified as per QC standards and is approved.


Name: Dr. Praveen Kumar
Department: Quality Control
Designation: QC Head