

**MOLBIO DIAGNOSTICS LIMITED**  
**EN ISO 13485:2016 certified company**  
**CERTIFICATE OF ANALYSIS**

Licence No.: MFG/IVD/2019/000018

Date: 15/06/2026

Document ID: COA/TN/I0001AB

|                   |                                                                            |
|-------------------|----------------------------------------------------------------------------|
| Product Name      | Truenat <sup>®</sup> Influenza A/B                                         |
| Generic Name      | Chip-based Real Time Duplex PCR Test for Influenza A and Influenza B Virus |
| Lot No.           | I0001AB                                                                    |
| Set No.           | NA                                                                         |
| Mfg. Date         | 2026-04                                                                    |
| Exp. Date         | 2028-03                                                                    |
| Pack Size         | 5T, 25T                                                                    |
| Storage condition | 2°C to 30°C                                                                |

**PRECAUTION:** Refer Safety Data Sheet and Instructions for Use.

| Sr. No. | Components in each kit                                                                                       | Specification                          | Observation             |
|---------|--------------------------------------------------------------------------------------------------------------|----------------------------------------|-------------------------|
| 01      | Number of pouches as per pack size                                                                           | Should be present                      | Present                 |
| 02      | Pouches Contents: 1 chip, 1 micro tube, 1 desiccant pouch and 1 DNase and RNase free filter barrier microtip | Should be present                      | Present                 |
| 03      | Colour of desiccant pouch                                                                                    | Should be present and Orange in Colour | Present / Orange Colour |
| 04      | Package Insert                                                                                               | Should be present                      | Present                 |
| 05      | Seal integrity test                                                                                          | Should pass                            | Pass                    |

| Performance testing data |              |              |              |
|--------------------------|--------------|--------------|--------------|
| Samples                  | Ct Values    |              |              |
|                          | Influenza A  | Influenza B  | Control      |
| Non Template Control     | ND           | ND           | ND           |
| Negative Sample          | ND           | ND           | 28.55 ± 1.02 |
| Positive Sample 1        | 31.04 ± 0.53 | ND           | 29.19 ± 0.90 |
| Positive Sample 2        | ND           | 31.63 ± 1.25 | 28.61 ± 0.78 |
| Positive Sample 3        | 26.74 ± 0.33 | 27.46 ± 0.71 | 29.85 ± 0.68 |

ND – Not detected, Ct- Cycle threshold

Mfg- Manufacturing, Exp- Expiry

The lot is tested and verified as per QC Standards and is approved.

for Name: Dr. Praveen Kumar MK  
Department: Quality Control  
Designation: QC Head

