

MOLBIO DIAGNOSTICS LIMITED
EN ISO 13485:2016 certified company
CERTIFICATE OF ANALYSIS

Licence no.: MFG/IVD/2025/000148

Date: 24/06/2026
Document ID: COA/TN/PCV008

Product Name	Truenat [®] Positive Control Kit - Panel V
Generic Name	Control Kit for Truenat [®] Chip-based micro PCR Tests: H3N2, Zika, Leish, ScrubT, Cholera, and Shigella
Lot No.	PCV008
Mfg. Date	2026-06
Exp. Date	2027-05
Pack Size	8T
Storage condition	2°C to 40°C


PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
01	Negative control	Should be present	Present
02	Positive control (08- strip tubes containing dried down positive control)	Should be present	Present
03	Package insert	Should be present	Present

Performance testing data			
Sr. No.	Truenat® test name	Positive control	Negative control
1	H3N2	29.18 ± 0.21	ND
2	Zika	27.94 ± 0.71	ND
3	Leish	28.91 ± 0.81	ND
4	Scrub T	30.70 ± 0.48	ND
5	Cholera	29.56 ± 0.63	ND
6	Shigella	30.02 ± 0.27	ND

*Mfg- Manufacturing, Exp- Expiry

The lot is tested and verified as per QC standards and is approved.



for Name: Dr. Praveen Kumar MK
Department: Quality Control
Designation: QC Head