

MOLBIO DIAGNOSTICS LIMITED
EN ISO 13485:2016 certified company
CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2021/000064

Date: 11/07/2026

Doc ID: COA/TP/095USPKL

Product Name	Trueprep [®] AUTO Universal Cartridge Based Sample Prep Kit
Generic Name	Sample preparation kit for nucleic acid extraction on Trueprep [®] AUTO
Lot No.	095USPKL
Mfg. Date	2026-06
Exp. Date	2028-05
Pack Size	25T
Storage condition	2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep[®] AUTO Universal Cartridge Based Sample prep kit)	(Present (P)/Absent (A))	
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml Graduated)	Present	Present
4	Package Insert	Present	Present

Performance on Truelab® Real Time micro PCR Analyzers						
Sr. No.	Specificity Tests	Specification		Test Lot		
		Test	Control	Test 1 (Average)	Test 2 (Average)	Control (Average)
Negative Control						
1	Negative HCV	ND	Control should be detected	ND	NA	27.76
2	Negative HIV-1	ND		ND	NA	26.98
3	Negative MTB	ND		ND	NA	27.72
4	Negative MTB Plus	ND		ND	NA	27.80
Positive Control						
1	Positive MTB	Test and Control should be detected		31.30	NA	27.59
2	Positive HCV	Test and Control should be detected		31.99	NA	26.95
Non Template Control Test for COVID-19						
1	NTC	Test 1, Test 2 and Control should not be detected		ND	ND	ND

*Ct-Cycle Threshold
Template Control

*IC- Internal Control

*ND- Not Detected

*NA- Not Applicable

*NTC- Non

The lot is tested and verified as per QC standards and is approved.


Name: Dr. Praveen Kumar
Department: Quality Control
Designation: QC Head