

MOLBIO DIAGNOSTICS LIMITED
EN ISO 13485:2016 certified company
CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2025/000136

Date: 11/08/2026

Doc ID: COA/TP/014USPT

Product Name	Trueprep [®] AUTO Universal Sample Pre-treatment Pack
Generic Name	Universal Sample pre-treatment pack for Non MTB samples
Lot No.	014USPT
Mfg. Date	2026-07
Exp. Date	2028-06
Pack Size	50T
Storage condition	2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep[®] AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
3	Package Insert	Present	Present

Performance on Truelab® Real Time micro PCR Analyzer					
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test (Average)	IC (Average)
Negative Control					
1	Negative HBV	ND	Control should be detected	ND	26.52
2	Negative HCV	ND	Control should be detected	ND	27.17
Positive Control					
1	Positive HBV	Test and Control should be detected		30.70	27.92
2	Positive HCV	Test and Control should be detected		32.71	26.34
Non Template Control					
1	NTC	Test and Control should not be detected		ND	ND

*IC- Internal Control *ND- Not Detected *NTC- Non Template Control

The lot is tested and verified as per QC standards and is approved.


Name: Dr. Praveen Kumar
Department: Quality Control
Designation: QC Head

