

MOLBIO DIAGNOSTICS LIMITED
EN ISO 13485:2016 certified company
CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2025/000148

Date: 25/07/2026

Doc ID: COA/TP/ATM116

Product Name	Trueprep [®] AUTO Transport Medium for Swab Specimen Pack
Generic Name	Medium for Collection, Decontamination and Transport of Swab Specimen for Trueprep [®] AUTO
Lot No.	ATM116
Mfg. Date	2026-06
Exp. Date	2028-05
Pack Size	5T, 25T
Storage condition	2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep[®] AUTO Transport Medium for Swab Specimen Pack)	[Present (P)/Absent (A)]	
1	Transport Medium for Swab Specimen Tubes	Present	Present
2	Package Insert	Present	Present

Performance on Truelab® Real Time micro PCR Analyzers						
Sr. No.	Specificity Tests	Specification		Test Lot		
		Test	Control	Test 1 Orf1A (Average)	Test 2 Egene (Average)	Control (Average)
Negative Control						
1	Negative COVID-19	ND	Control should be detected	ND	ND	27.82
Positive Control						
1	Positive COVID-19	Test and Control should be detected		29.64	30.60	24.18
Non Template Control Test for COVID-19						
1	NTC	Test 1, Test 2 and Control should not be detected		ND	ND	ND

*ND- Not Detected

*NTC- Non Template Control

The lot is tested and verified as per QC Standards and is approved.



Name: Dr. Praveen Kumar MK

Department: Quality Control

Designation: QC Head

for