

MOLBIO DIAGNOSTICS LIMITED
EN ISO 13485:2016 certified company
CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2025/000136

Date: 01/09/2026

Doc ID: COA/TP/275MSPT

Product Name	Trueprep [®] AUTO MTB Sample Pre-treatment Pack
Generic Name	Sample pre-treatment pack for MTB samples
Lot No.	275MSPT
Mfg. Date	2026-08
Exp. Date	2028-07
Pack Size	50T
Storage condition	2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep[®] AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present

Performance on Truelab® Real Time micro PCR Analyzer						
Sr. No.	Specificity Tests	Specification		Test Lot		
		Target	Control	Target 1 (Average)	Target 2 (Average)	Control (Average)
Negative Control						
1	Negative MTB	ND	Control should be detected	ND	NA	26.47
Positive Control						
1	Positive MTB	Target and Control should be detected		27.24	NA	27.31
2	Positive MTB	Target and Control should be detected		32.30	NA	27.03
Non Template Control Test for MTB						
1	NTC	Target 1, Target 2 and Control should not be detected		ND	ND	ND

*ND- Not Detected *NA- Not Applicable *NTC- Non Template Control

The lot is tested and verified as per QC standards and is approved.


Name: Dr. Praveen Kumar M.K.
Department: Quality Control
Designation: QC Head

