molbio®
INNOVATE • IMPACT • INSPIRE

**Molbio Diagnostics Limited**
(Formerly Molbio Diagnostics Private Limited)
L-46, Phase II D, Verna Industrial Estate
Verna, Goa - 403722, India
+91 832 6724888 www.molbiodiagnostics.com

## MOLBIO DIAGNOSTICS LIMITED

### EN ISO 13485:2016 certified company

### CERTIFICATE OF ANALYSIS

**License No.:** MFG/IVD/2025/000148

**Date:** 29/10/2026

**Doc ID:**COA/TP/MSPT333

| | |
|---|---|
| **Product Name** | Trueprep® AUTO MTB Sample Pre-treatment Pack |
| **Generic Name** | Sample pre-treament pack for MTB samples |
| **Lot No.** | MSPT333 |
| **Mfg. Date** | 2026-10 |
| **Exp. Date** | 2028-09 |
| **Pack Size** | 25T |
| **Storage condition** | 2°C to 40°C |

**PRECAUTION:** Refer Safety Data Sheet and follow instructions for use.

| Sr. No. | Components in each kit | Specification | Observation |
|---|---|---|---|
| **A** | **Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)** | **[Present (P)/Absent (A)]** | |
| 1 | Lysis Buffer | Present | Present |
| 2 | Liquefaction Buffer | Present | Present |
| 3 | Disposable Transfer Pipette (1 ml Graduated) | Present | Present |
| 4 | Package Insert | Present | Present |
| **B** | **Leak Test** | **Specification** | **Observation** |
| 1 | Leak Test | Should Pass | Pass |



sales@molbiodiagnostics.com / admin@molbiodiagnostics.com | **CIN:** L33125GA2000PLC002909 | **GSTIN:** 30AACCM8275E1ZO

molbio®
INNOVATE • IMPACT • INSPIRE

**Molbio Diagnostics Limited**
(Formerly Molbio Diagnostics Private Limited)
L-46, Phase II D, Verna Industrial Estate
Verna, Goa - 403722, India
+91 832 6724888 www.molbiodiagnostics.com

| Performance on Truelab® Real Time micro PCR Analyzer | | | | | |
|---|---|---|---|---|---|
| Sr. No. | Specificity Tests | Specification | | Test Lot | |
| | | Target | Control | Target (Average) | Control (Average) |
| **Negative Control** | | | | | |
| 1 | Negative MTB | ND | Control should be detected | ND | 27.82 |
| **Positive Control** | | | | | |
| 1 | Positive MTB | Target and Control should be detected | | 28.68 | 27.20 |
| 2 | Positive MTB | Target and Control should be detected | | 24.12 | 27.12 |
| **Non Template Control** | | | | | |
| 1 | NTC | Target and Control should not be detected | | ND | ND |

*ND- Not Detected *NTC- Non Template Control

**The lot is tested and verified as per QC Standards and is approved.**

MOLBIO DIAGNOSTICS LIMITED GOA

**Name:** Dr. Praveen Kumar M.K.
**Department:** Quality Control
for **Designation:** QC Head



sales@molbiodiagnostics.com / admin@molbiodiagnostics.com | **CIN:** L33125GA2000PLC002909 | **GSTIN:** 30AACCM8275E1ZO