molbio®
INNOVATE • IMPACT • INSPIRE

Molbio Diagnostics Limited
(Formerly Molbio Diagnostics Private Limited)
L-46, Phase II D, Verna Industrial Estate
Verna, Goa - 403722, India
+91 832 6724888 www.molbiodiagnostics.com

MOLBIO DIAGNOSTICS LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Licence no.: MFG/IVD/2025/000148

Date: 29/10/2026
Document ID: COA/TN/PCI072

Product Name	**Truenat® Positive Control Kit - Panel I**
Generic Name	**Control Kit for Truenat® Chip-based micro PCR Tests: MTB, MTB Plus, MTB-RIF Dx, Beta CoV, SARS CoV-2, COVID-19 and H1N1**
Lot No.	**PCI072**
Mfg. Date	**2026-10**
Exp. Date	**2027-09**
Pack Size	**8T**
Storage condition	**2°C to 40°C**

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
01	Negative control	Should be present	Present
02	Positive control (08- strip tubes containing dried down positive control)	Should be present	Present
03	Package insert	Should be present	Present

sales@molbiodiagnostics.com / admin@molbiodiagnostics.com | **CIN:** L33125GA2000PLC002909 | **GSTIN:** 30AACCM8275E1ZO

molbio®
INNOVATE • IMPACT • INSPIRE

Molbio Diagnostics Limited
(Formerly Molbio Diagnostics Private Limited)
L-46, Phase II D, Verna Industrial Estate
Verna, Goa - 403722, India
+91 832 6724888 www.molbiodiagnostics.com

Performance testing data			
Sr. No.	**Truenat® test name**	**Positive control**	**Negative control**
1	MTB	29.37 ± 0.74	ND
2	MTB Plus	29.42 ± 0.89	ND
3	MTB-RIF Dx	RIF Resistance detected	Indeterminate
4	SARS CoV-2	27.52 ± 0.56	ND
5	COVID-19	Orf 1A: 27.69 ± 0.20 E gene: 29.49 ± 0.32	ND
6	H1N1	27.25 ± 0.40	ND

*Mfg- Manufacturing, Exp- Expiry

The lot is tested and verified as per QC standards and is approved.

MOLBIO DIAGNOSTICS LIMITED GOA

Name: Dr. Praveen Kumar M.K.
Department: Quality Control
Designation: QC Head

sales@molbiodiagnostics.com / admin@molbiodiagnostics.com | **CIN:** L33125GA2000PLC002909 | **GSTIN:** 30AACCM8275E1ZO