molbio®
INNOVATE • IMPACT • INSPIRE

Molbio Diagnostics Limited
(Formerly Molbio Diagnostics Private Limited)
L-46, Phase II D, Verna Industrial Estate
Verna, Goa - 403722, India
+91 832 6724888 www.molbiodiagnostics.com

CERTIFICATE OF ANALYSIS

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack
Generic Name	Sample pre-treatment pack for MTB samples
License No.	MFG/IVD/2025/000136
Lot No.	297MSPT
Mfg. Date	2026-11
Exp. Date	2028-10
Catalogue No./(Pack Size)	60204AS50/(50T)
Storage condition	2°C to 40°C
Document ID	COA/TP/297MSPT
Date of Release	10/11/2026

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Physical Verification data			
Sr. No.	**Components in each kit**	**Specification**	**Observation**
A	**Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)**	**[Present (P)/Absent (A)]**	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present
B	**Leak Test**	**Specification**	**Observation**
1	Leak test	Should Pass	Pass

sales@molbiodiagnostics.com / admin@molbiodiagnostics.com | **CIN:** L33125GA2000PLC002909 | **GSTIN:** 30AACCM8275E1ZO

molbio®
INNOVATE • IMPACT • INSPIRE

Molbio Diagnostics Limited
(Formerly Molbio Diagnostics Private Limited)
L-46, Phase II D, Verna Industrial Estate
Verna, Goa - 403722, India
+91 832 6724888 www.molbiodiagnostics.com

Performance testing data						
Sr. No.	Specificity Tests	Specification		Test Lot		
		Target	Control	Target 1 (Average)	Target 2 (Average)	Control (Average)
Negative Control						
1	Negative MTB	ND	Control should be detected	ND	NA	26.39
Positive Control						
1	Positive MTB	Target and Control should be detected		26.21	NA	27.63
2	Positive MTB	Target and Control should be detected		31.01	NA	27.56
Non Template Control						
1	NTC	Target and Control should not be detected		ND	ND	ND

Mfg- Manufacturing, Exp- Expiry, No.- Number, ID- Identification

The lot is tested and verified as per QC standards and is approved.

MOLBIO DIAGNOSTICS LIMITED GOA

Name: Dr. Praveen Kumar M.K.
Department: Quality Control
Designation: QC Head

sales@molbiodiagnostics.com / admin@molbiodiagnostics.com | **CIN:** L33125GA2000PLC002909 | **GSTIN:** 30AACCM8275E1ZO