molbio®
INNOVATE • IMPACT • INSPIRE

Molbio Diagnostics Limited
(Formerly Molbio Diagnostics Private Limited)
L-46, Phase II D, Verna Industrial Estate
Verna, Goa - 403722, India
+91 832 6724888 www.molbiodiagnostics.com

CERTIFICATE OF ANALYSIS

Product Name	Trueprep® AUTO Transport Medium for Swab Specimen Pack
Generic Name	Medium for Collection, Decontamination and Transport of Swab Specimen for Trueprep® AUTO
License No.	MFG/IVD/2025/000148
Lot No.	ATM120
Mfg. Date	2026-11
Exp. Date	2028-10
Catalogue No./(Pack Size)	60206TS05/(5T), 60206TS20/(20T), 60206TS25/(25T), 60206TS50/(50T)
Storage condition	2°C to 40°C
Document ID	COA/TP/ATM120
Date of Release	11/11/2026

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Physical Verification Data			
Sr. No.	**Components in each kit**	**Specification**	**Observation**
A	**Contents in the carton (Trueprep® AUTO Transport Medium for Swab Specimen Pack)**	**[Present (P)/Absent (A)]**	
1	Transport Medium for Swab Specimen Tubes	Present	Present
2	Package Insert	Present	Present
B	**Leak Test**	**Specification**	**Observation**
1	Leak Test	Should Pass	Pass

sales@molbiodiagnostics.com / admin@molbiodiagnostics.com | **CIN:** L33125GA2000PLC002909 | **GSTIN:** 30AACCM8275E1ZO

molbio®
INNOVATE • IMPACT • INSPIRE

Molbio Diagnostics Limited
(Formerly Molbio Diagnostics Private Limited)
L-46, Phase II D, Verna Industrial Estate
Verna, Goa - 403722, India
+91 832 6724888 www.molbiodiagnostics.com

Performance on Truelab® Real Time micro PCR Analyzers						
Sr. No.	Specificity Tests	Specification		Test Lot		
		Target	Control	Target 1 E.gene (Average)	Target 2 Orf1A (Average)	Control (Average)
Negative Control						
1	Negative COVID-19	ND	Control should be detected	ND	ND	24.86
Positive Control						
1	Positive COVID-19	Target and Control should be detected		23.19	22.89	24.20
Non Template Control Test for COVID-19						
1	NTC	Target 1, Target 2 and Control should not be detected		ND	ND	ND

Mfg- Manufacturing, Exp- Expiry, No.- Number, ID- Identification, ND- Not Detected, NTC- Non Template Control

The lot is tested and verified as per QC standards and is approved.

11/11/26

MOLBIO DIAGNOSTICS LIMITED GOA

for **Name:** Dr. Praveen Kumar M.K.
Department: Quality Control
Designation: QC Head

sales@molbiodiagnostics.com / admin@molbiodiagnostics.com | **CIN:** L33125GA2000PLC002909 | **GSTIN:** 30AACCM8275E1ZO