Molbio Diagnostics Limited
(Formerly Molbio Diagnostics Private Limited)
L-46, Phase II D, Verna Industrial Estate
Verna, Goa - 403722, India
+91 832 6724888 www.molbiodiagnostics.com

CERTIFICATE OF ANALYSIS

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack
Generic Name	Sample pre-treament pack for MTB samples
License No.	MFG/IVD/2025/000148
Lot No.	MSPT334
Mfg. Date	2026-11
Exp. Date	2028-10
Catalogue No./(Pack Size)	60204AS20/(20T), 60204AS50/(50T)
Storage condition	2°C to 40°C
Document ID	COA/TP/MSPT334
Date of Release	07/11/2026

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Physical Verification Data			
Sr. No.	**Components in each kit**	**Specification**	**Observation**
A	**Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)**	**[Present (P)/Absent (A)]**	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present
B	**Leak Test**	**Specification**	**Observation**
1	Leak Test	Should Pass	Pass

sales@molbiodiagnostics.com / admin@molbiodiagnostics.com | **CIN:** L33125GA2000PLC002909 | **GSTIN:** 30AACCM8275E1ZO

Molbio Diagnostics Limited
(Formerly Molbio Diagnostics Private Limited)
L-46, Phase II D, Verna Industrial Estate
Verna, Goa - 403722, India
+91 832 6724888 www.molbiodiagnostics.com

Performance on Truelab® Real Time micro PCR Analyzer					
Sr. No.	Specificity Tests	Specification		Test Lot	
		Target	Control	Target (Average)	Control (Average)
Negative Control					
1	Negative MTB	ND	Control should be detected	ND	27.33
Positive Control					
1	Positive MTB	Target and Control should be detected		29.29	28.01
2	Positive MTB	Target and Control should be detected		24.65	27.11
Non Template Control					
1	NTC	Target and Control should not be detected		ND	ND

Mfg- Manufacturing, Exp- Expiry, No.- Number, ID- Identification

The lot is tested and verified as per QC standards and is approved.

MOLBIO DIAGNOSTICS LIMITED GOA

04/11/26

for **Name:** Dr. Praveen Kumar M.K.
Department: Quality Control
Designation: QC Head