Molbio Diagnostics Limited
(Formerly Molbio Diagnostics Private Limited)
L-46, Phase II D, Verna Industrial Estate
Verna, Goa - 403722, India
+91 832 6724888 www.molbiodiagnostics.com

CERTIFICATE OF ANALYSIS

Product Name	Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit
Generic Name	Sample preparation kit for nucleic acid extraction on Trueprep® AUTO v2
Licence No.	MFG/IVD/2019/000003
Lot No.	VUCPK564L
Mfg. Date	2026-10
Exp. Date	2028-09
Catalogue No./(Pack size)	60207AR50 /(50T)
Storage condition	2°C to 40°C
Document ID	COA/TP/VUCPK564L
Date of Release	09/11/2026

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	**Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)**	**[Present (P)/Absent (A)]**	
1	Universal Reagent Pack	Should be Present	P
2	Universal Cartridge Pack	Should be Present	P
3	Disposable Transfer Pipette (3 ml Graduated)	Should be Present	P
4	Package Insert	Should be Present	P
B	**Pouch Seal Integrity**	**Specification**	**Observation**
1	Pouch Seal Integrity Test	Should Pass	Pass

sales@molbiodiagnostics.com / admin@molbiodiagnostics.com | **CIN:** L33125GA2000PLC002909 | **GSTIN:** 30AACCM8275E1ZO

molbio®
INNOVATE • IMPACT • INSPIRE

Molbio Diagnostics Limited
(Formerly Molbio Diagnostics Private Limited)
L-46, Phase II D, Verna Industrial Estate
Verna, Goa - 403722, India
+91 832 6724888 www.molbiodiagnostics.com

Performance on Truelab® Real Time micro PCR Analyzer						
Sr. No.	Specificity Tests	Specification		Test Lot		
		Target	Control	Target 1 (Average)	Target 2 (Average)	Control (Average)
Negative Control						
1	Negative MTB	ND	Control should be detected	ND	NA	26.13
2	Negative HCV	ND		ND	NA	26.22
3	Negative HIV-1/ HIV-2	ND		ND	ND	28.10
4	Negative MTB Plus	ND		ND	NA	27.25
Positive Control						
1	Positive MTB	Target and Control should be detected		30.26	NA	26.73
2	Positive HCV	Target and Control should be detected		31.77	NA	26.82
3	Positive MTB Plus	Target and Control should be detected		28.36	NA	27.84
Non Template Control Test for COVID-19						
1	NTC	Target 1, Target 2 and Control should not be detected		ND	ND	ND

Mfg- Manufacturing, Exp- Expiry, No.- Number, ID- Identification, ND- Not Detected, NA- Not Applicable, NTC- Non Template Control

The lot is tested and verified as per QC standards and is approved.

MOLBIO DIAGNOSTICS LIMITED GOA

for 04/11/26

Name: Dr. Praveen Kumar M.K.
Department: Quality Control
Designation: QC Head

sales@molbiodiagnostics.com / admin@molbiodiagnostics.com | **CIN:** L33125GA2000PLC002909 | **GSTIN:** 30AACCM8275E1ZO